DECLARATION OF CONFORMITY

PolarSeal

DECLARATION OF CONFORMITY

This declaration of conformity is issued under the sole responsibility of PolarSeal Tapes and Conversions Ltd in compliance to Article 19 of EU MDR 2017/745. We hereby declare that the medical device specified below meets the provision of the Annex IV of Regulation EU MDR 2017/745 for medical devices.

The declaration is based on:

Manufacturer	Polarseal Tapes and Conversions Ltd		
Manufacturer Address	Guildford Road Trading Estate,		
	Farnham, Surrey GU9 9PZ		
SRN (Single Registration	NL-AR-00000116		
Number)	THE FIRST CONTROL OF THE FIRST		
Authorized	EMERGO EUROPE		
Representative Name			
	Prinsessegracht 20, 2514 AP		
Authorized	The Hague,		
Representative Address	The Netherlands		
Name of the Device	Ectovise C-Smile T1		
Product code	CS-T1		
Intended Use	The C-Smile T1 transparent face mask is a non-sterile medical device and recommended for single shift use. Its primary purpose is for healthcare		
Intellueu Ose			
	professionals, healthcare workers, first responders and other personnel to protect		
	the patient from infection via direct breath, vapour, liquid splashes and droplets at		
	the same time as letting their facial expressions being seen.		
Classification with rule	C-Smile T1 is classified as a Class I non-invasive, single-use, non-sterile device		
number	-,		
Conformity assessment	Medical Device Regulation 2017/745 in accordance with Rule 1 Annex VIII.		
route			
Harmonized Standards	Meets the splash resistance performance requirements for type IIR surgical masks		
	(EN14683:2019+AC:2019, chapter 5.2.7).		
	Meets the microbial cleanliness (bioburden) performance requirements for type IIR		
941	surgical masks (EN14683:2019+AC:2019, chapter 5.2.7).		
	Meets the breathability performance requirements for type IIR surgical masks		
	(EN14683:2019+AC:2019, chapter 5.2.7).		
	Meets the bacterial filtration efficiency performance requirements for type IIR		
	surgical masks (EN14683:2019+AC:2019, chapter 5.2.7).		
	Meets the EN166:2002 resistance to fogging (anti-fog) performance requirement		
Start of CE marking	February 2022		

Name	Position	Signature	Date (dd/mm/yyyy)
204N BICH	DIRECTOR	all or	11/02/2022